OPERATING MANUAL

EchoTherm™
DIGITAL HOT PLATE
MODEL HP30 AND HP30A

DOCUMENT NUMBER HP30-03
Revised 10 December 2011

TORREY PINES SCIENTIFIC, INC.
2713 Loker Ave. West
Carlsbad, CA 92010

TELEPHONE: (760)-930-9400
TOLL FREE: (866)-573-9104
FAX: (760)-930-9480
E-MAIL: info@torreypinesscientific.com
WEB SITE: www.torreypinesscientific.com

TABLE OF CONTENTS

		PAGE
I.	INTRODUCTION	3
II.	WARRANTY	3
III.	RETURN OF ITEMS	3
IV.	LABELS AND SYMBOLS	3
V.	CAUTIONS	4
VI.	GENERAL DESCRIPTION	5
VII.	FRONT AND REAR PANEL CONTROLS	6, 7
VIII.	SET-UP INSTRUCTIONS	8
IX.	DISPLAY AND KEYBOARD DESCRIPTION	9, 10
X.	SETTING TEMPERATURE, RAMP RATE, TIMER, AND AUTO-OFF	11, 12
XI.	TEMPERATURE CALIBRATION	13, 14
XII	POWER INTERRUPT PROTECTION	14
XIII.	RS232 INTERFACE	15, 16
XIV.	CLEANING AND MAINTENANCE	16

I. INTRODUCTION

Congratulations on your purchase of a Torrey Pines Scientific EchoTherm™ Digital Hot Plate. Please read the instructions carefully to insure that you receive the maximum benefit from it. Also, be sure to fill out and return the enclosed warranty registration card.

II. WARRANTY

Torrey Pines Scientific warrants this product to be free from defects in material and workmanship for a period of one year from the date of purchase. If repair or adjustment is necessary and has not been the result of abuse or misuse within the one year period, please return--freight prepaid-- and correction of the defect will be made without charge.

Out-of-warranty products will be repaired on a charge basis.

III. RETURN OF ITEMS

Authorization must be obtained from our Customer Service Department before returning items for any reason! When applying for authorization, please include data regarding the reason the items are to be returned. For your protection, items must be carefully packed to prevent damage in shipment and insured against possible damage or loss. Torrey Pines Scientific will not be responsible for damage resulting from careless or insufficient packing. A 15% restocking charge will be made on all unauthorized returns.

NOTE: Torrey Pines Scientific reserves the right to make improvements in design, construction and appearance without notice.

IV. LABELS AND SYMBOLS

There are various labels on the body of this unit. Listed below are the labels and their meanings.

This symbol means" ATTENTION, THE INSTRUCTION MANUAL IS TO BE CONSULTED FOR FURTHER INFORMATION".

This symbol means "WARNING, HOT SURFACE".

This symbol means "GROUND or EARTH CONNECTION".

V. CAUTIONS

HEATER PLATE SURFACE

The Torrey Pines Scientific HP30 Hot Plate is capable of temperatures in excess of 450°C at the plate surface. Touching the heated surface will cause severe burns. **USE EXTREME CAUTION AT ALL TIMES.** Never leave your hot plate accessible to others while it is hot. Although the unit is equipped with a "HOT WARNING" indicator on the front panel, do not rely on this alone. Never touch the heating surface unless you are absolutely sure that it is cool.
NOTE: DO NOT INSTALL THIS UNIT CLOSER THAN SIX (6) INCHES (15.24 CM) TO A WALL OF COMBUSTIBLE MATERIAL. ALSO, THIS UNIT IS NOT FOR USE WITH FLAMMABLE LIQUIDS.

TEMPERATURE PROBE

When attempting to control PROBE TEMPERATURE, it is necessary to plug in the temperature probe and to place it in the sample AT ALL TIMES. If the probe is not placed into the sample and plugged into the rear of the hot plate, the unit will not be able to sense the rising temperature of the sample as heat is applied. This will result in driving the heater to its maximum and could result in ruining the sample.

Temperature probes vary in size and material. The most commonly used probe is the stainless steel 6-inch immersion probe (part number HS30-600). For other probes available in other lengths and materials, call the factory or consult your price list. All probes come with 3-foot (91.44 cm) interconnect cable and are 3/16" (47.6 mm) in diameter.

Stainless steel probes work well with all but the most aggressive chemicals. Solid Teflon will work with the most aggressive chemicals, but only to about 260°C. Glass probes are available for use with aggressive chemicals at higher temperatures.

ELECTRICAL

These hot plates are made in models that operate at 100, 115, and 230 volts AC. Be certain that your voltage matches the unit you receive. Check the name plate on the bottom for the voltage setting on your unit. Take the normal care and precaution one would use with any electrical appliance. Be very careful to keep the AC line cord away from the hot plate. **NOTE: ALL FUSES ARE FAST BLOW. THEY ARE RATED AT 8 AMPS, 125 VOLTS FOR CERAMIC TOPPED 100 AND 115 VAC UNITS, AND 4 AMPS, 250 VOLTS FOR 230 VAC UNITS. CAST ALUMINUM TOPPED UNITS ARE 10 AMPS, 125 VOLTS FOR 100 AND 115 VAC UNITS AND 5 AMPS, 250 VOLTS FOR 230 VAC UNITS.**

VI. GENERAL DESCRIPTION

The Torrey Pines Scientific HP30 Series Hot Plate is a general purpose, digital hot plate. It is available with a milled-flat cast aluminum top or a solid ceramic top. Both are 8" x 8" (20.32 cm x 20.32 cm) in size. All functions are settable from the digital front panel and display. These units when received display temperature in degrees centigrade. They can be switched to read in degrees Fahrenheit. See page 6.

HEATER

Either the plate surface temperature or the actual sample temperature may be set by the user. A sensor in the plate is used to monitor surface temperature, or, alternately, the 100-ohm, 3-wire, platinum RTD temperature probe may be connected to the rear of the unit and inserted into the sample to measure sample temperature. These probes are available in stainless steel, solid Teflon or borosilicate glass. When a temperature is set by the user, power is applied to the heater to precisely control the temperature at the plate surface or at the sample, as directed by the user.

An optional “ramp value” may be entered which causes the temperature to approach the target value at a controlled rate of change. This can be either an increase or decrease in temperature. Ramping temperature is always displayed in centigrade.

TIMER

The HP30 comes with a count down timer that is settable and displays in hours, minutes, and seconds. The timer can be set to 99:59:59 hours. When the timer is set it starts to count down to zero. At zero the timer will sound an audible alarm that will ring ten times. In addition to the timer, the unit has an AUTO-OFF function. When the AUTO-OFF function is activated the heater will turn off at the end of any count down timer setting.

ALARMS

In addition to the audible alarm associated with the count down timer, the HP30 has an **over temperature alarm** that activates at 455°C**.** If, for any reason, the set temperature runs away and exceeds 455°C (852ºF), the HP-30 will turn on the over temperature alarm LED, sound an audible alarm, and shut the heater off. If or when the heater plate surface cools to below 450°C, the heater will turn on again, and the alarm LED and audible alarm will turn off. In addition, all circuitry in the unit is backed up by a separate watchdog circuit to assure safety.

NOTE: Failing to place the temperature probe in the solution and then setting a probe temperature will cause the heater to heat to over 455°C and set off the OVER TEMPERATURE ALARM. Also, the OVER TEMPERATURE ALARM may come on if a very large sample is placed on the unit and a probe temperature is set. This setting can cause the heater surface to go over 455°C to achieve the probe temperature set to heat the large sample.

VII. FRONT AND REAR PANEL CONTROLS

FRONT PANEL

Shown above are the keyboard and display for the HP30 Digital Hot Plate.

The front panel has a tactile touch membrane keyboard with audible feedback. The keyboard is used to set all operating parameters. The display is an LCD type made up of four rows of numbers and several icons which function to display all parameters when the unit is running and during setting. When the unit is turned on, the display will come on and show the actual plate surface temperature. If a probe is used, it will display the probe temperature as well. The timer and ramp functions will show zeros.

There are two LED's on the front panel as well. One is actuated when the plate surface goes over 50°C to remind the user that the plate surface is hot enough to cause burns. The second is for the over temperature alarm.

To change the unit from reading in ºC to reading in ºF or vise-versa hold the HEAT OFF button down for 3 seconds. The display will switch over. Degrees F is indicated when the "C" symbol next to the temperature reading for both plate and probe is missing. *The RAMP value is always set in ºC/hour via the keypad or RS232 I/O port.*

REAR PANEL

The AC power connector jack, fuse holder and switch are a module mounted on the left rear of the unit. The temperature probe jack is mounted to the right. Note the polarity on the probe jack. Be sure not to force this connection when plugging in the temperature probe.

The AC power jack is a three-prong, male plug combining the snap-in fuse holder and AC switch. The fuse types used appear on the label on the rear of the unit. Both neutral and high lines are fused.

CAUTION: If the fuses blow repeatedly, contact your dealer or the manufacturer.

The temperature sensor jack is a 5-pin DIN jack which is used with a 100-ohm platinum RTD temperature probe.

VIII. SET UP PARAMETERS AND INSTRUCTIONS

The HP30 Digital Hot Plate is a very simple instrument to use. Follow the instructions below and you will agree.

SET UP PARAMETERS

1. Ambient operating temperature range is from 5°C to 40°C.

2. Maximum altitude of operation should not exceed 2000 meters.

3. Maximum ambient operating relative humidity should not exceed 80% at 31°C decreasing linearly to 50% relative humidity at 40°C.

SET UP INSTRUCTIONS

1. Place the unit on a level, dry bench or surface.

2. Plug the unit into a properly grounded, three-wire outlet of proper voltage.

3. Plug the temperature probe (if used) into the jack on the rear of the unit.

4. Place the sample on the heater plate and put the temperature probe (if used) into the sample container.

5. Turn on the unit by the switch on the rear panel. The unit will beep once and the display will light up. At this point the user can set or view any of the parameters of the unit.

6. Set target temperatures, timer, and ramp rate according to the instructions given in the following sections.

ENVIRONMENTAL INFORMATION

1. This unit is for installation category 2.

2. This unit is rated pollution degree 2.

IX. DISPLAY AND KEYBOARD DESCRIPTIONS

DISPLAY

The display has four rows of numbers and icons. From the top they are as follows:

Probe Temperature: The icon to the left of the numbers represents a probe in a solution. This row of numbers shows the probe temperature. These numbers will always be followed with a “C” which denotes that the temperature reading is in degrees centigrade unless the unit has been switched to read in degrees F. In that case the “C” next to the temperature display will not illuminate. The small “T” and “A” following that tell the user that the numbers displayed are the actual temperature of the probe in the solution or the target temperature as set by the user. These two letters will toggle between TARGET(T) and ACTUAL(A), and the numbers in the display will change accordingly to show the progress of a set temperature toward its target. If no target temperature is set, then the display will always show the actual probe temperature.

Plate Temperature: The icon to the left of the second row of numbers represents the plate temperature at its surface as measured with a sensor in the heater plate. The row of numbers to the right of the icon shows the plate surface temperature. These numbers also are followed by a “C” which denotes that the temperature reading is in degrees centigrade unless the display has been awitched to read in degrees F. In that case the “C” will not illuminate. The small “T” and “A” following that tell the user that the numbers displayed are the TARGET(T) or ACTUAL(A) plate surface temperatures. If no plate surface target temperature is set, the “T” and “A” will not toggle. However, the actual plate temperature will be displayed even when a probe temperature has been set.

Ramp: The icon next in line down the display looks like a staircase and is the RAMP icon. The number value set in the display is always in degrees centigrade per hour regardless of whether the temperature display is set to read in degrees C or F. This tells the user that the target temperature as set for probe or plate is moving from its starting point to the target temperature at a selected rate in degrees centigrade per hour.

Timer: The next icon down is the hourglass icon that represents the timer function. The numbers that can be set and displayed next to the icon are in hours, minutes, and seconds. The timer can be set to 99:59:59 maximum. The timer will count down to zero from its setting and sound an audible alarm and then start to count up letting the user know how long it has been since the timer counted down to zero. The words “AUTO OFF” next to the numbers are visible only when the AUTO-OFF function is actuated. When this function is actuated, the timer will count down to zero, sound an audible alarm, and turn the heater off.

KEYBOARD

Up/Down Arrows: The UP and DOWN arrows in the center of the keyboard are used to set values for the probe temperature, plate temperature, ramp, and timer. Pushing and holding the UP arrow will cause the display parameter as selected to scroll up in value. Pushing and holding the Down arrow will cause the display parameter as selected to go down in value. In addition, the DOWN arrow is used as "cancel" to turn off the beep when the timer counts down.

Probe Temperature Icon: The probe temperature icon is the one that matches the icon to the left of the top line of the display. Touching and releasing this icon selects the UP and DOWN arrows to set the PROBE TEMPERATURE.

Plate Temperature Icon: The plate temperature icon is the one that matches the second line down from the top in the display. Touching and releasing this icon selects the UP and DOWN arrows to set the PLATE TEMPERATURE.

Ramp Icon: The ramp icon is the one that matches the third line down from the top of the display. Touching and releasing this icon selects the UP and DOWN arrows to set the RAMP value.

Timer Icon: The timer icon is the one that matches the last line at the bottom of the display. Touching and releasing this icon selects the UP and DOWN arrows to set the TIMER value.

Auto-Off Key: Touching and releasing this key activates the AUTO-OFF function. When activated, this will shut off the heater when the timer counts to zero. This function can be turned on or off by touching the AUTO-OFF button at any time during a timed event. When activated, the words "AUTO-OFF" will be illuminated on the display.

Heater Off Key: Touching and releasing the HEATER OFF key will turn off the heater. Any instruction given the heater will now be lost and the heater target will return to zero. The HEATER OFF key also is used to change the unit temperature display from ºC to ºF and vise-versa. Holding the key down for 3 seconds will cause the display to change.

X. SETTING TEMPERATURE, RAMP RATE, TIMER AND AUTO-OFF

TEMPERATURE

The heater may be set to control either the plate surface or the sample itself. This is done by entering a "target" temperature from the front panel keyboard. The control electronics in the hot plate will automatically apply power to the heater plate to reach the desired temperature. The user may enter either a target probe temperature or a target plate temperature. Only one target temperature is allowed at one time. Setting a target probe temperature will erase any target plate temperature previously set and vice versa. Target temperatures may be set anywhere in the range from ambient to 400°C (752ºF) on the aluminum topped heater and from ambient to 450°C (842ºF) in the solid ceramic topped unit. To switch the unit to read in degrees F see page 6.

Setting Solution Temperature: To set a solution temperature, touch and release the solution temperature icon. The icon will start to flash on the display. Next, press the UP arrow until the display (top row) shows the temperature wanted. If the target is overshot while setting, push the DOWN arrow until the correct value wanted is in the display. The unit will count five seconds and stop flashing the icon and turn the heater on and drive the solution to the target temperature. Note that the solution temperature on the display will toggle between the TARGET (T) and ACTUAL (A) temperatures. When the target temperature is reached the "T" and "A" will alternate but the numerical value displayed will remain the same.

CAUTION: BE CERTAIN THE TEMPERATURE PROBE IS IN THE SOLUTION AND PLUGGED INTO THE REAR PANEL JACK WHEN A SOLUTION TEMPERATURE TARGET HAS BEEN ENTERED. FAILURE TO DO SO COULD DAMAGE YOUR SAMPLE BECAUSE THE HOT PLATE WILL DRIVE TO MAXIMUM WHILE SEEKING A TEMPERATURE IT CANNOT FIND.

Setting Plate Temperature: To set a plate temperature, touch and release the plate temperature icon. The plate icon will flash on the display. Next, press the UP arrow until the display (second row) shows the temperature wanted. If the target is overshot while setting, push the DOWN arrow until the correct value is in the display. The unit will count five seconds and the icon will stop flashing. The unit will now turn on the heater and drive the plate to the target temperature selected. Note that the display will toggle between the TARGET (T) and ACTUAL (A) temperature. Again, when the target temperature is reached, the "T" and "A" will alternate but the numerical value displayed will remain the same.

Note: When a heater value is turned off, the value is saved so that when the heater is next used that previously set value is displayed and set. To change that value use the UP or DOWN arrows.

Setting Ramp: **NOTE: IF A RAMP VALUE IS TO BE USED, THE RAMP VALUE MUST ALWAYS BE SET BEFORE THE TARGET TEMPERATURE IS SET. IT IS ALWAYS SET IN ºC only, even if the temperature display is reading in degrees F.** The ramp rate is settable from 1ºC/Hour to 450ºC/hour**.** Note that the unit will heat fastest when not using a ramp rate setting, even 450ºC/Hour. The ramp rate should only be used to slow the rate of heating.

To set a ramp value, touch and release the ramp icon. The display will now flash that icon. Press the UP arrow until the display (third row) shows the value desired. Remember, the ramp value is always in degrees centigrade per hour. The unit will count five seconds and stop flashing the icon. The heater will now go from its current value to the target temperature, either plate or solution, at the rate just instructed.

NOTE: It is valuable to note that the ramp can be set to go up to a target temperature or down from a higher temperature to a lower one. However, the unit can never ramp down any faster than the sample can cool naturally. When ramping down, the heater is used to slow a cooling rate only.

Setting the Timer: To set a timer value, touch and release the timer icon (hourglass). That icon will now flash on the display. Note that the seconds digits in the display will flash. Seconds can now be set if wanted BY PRESSING THE up arrow. To set minutes, press the timer icon again. Note that the minutes digits will now flash. Minutes can now be set by touching the UP arrow. To set hours, press the timer icon again and the hours digits will flash. Hours may now be set by pressing the UP arrow. To set any of these above, press the UP arrow until the display reaches the desired value. If only seconds or minutes are set, continue to press the timer icon until it scrolls past hours. Touching the timer icon again will start the timer.

Setting Auto-Off: To set the auto-off function, touch and release the AUTO-OFF button. The display will illuminate the words “AUTO-OFF” to the right of the timer. This tells the user that the auto-off function is set. To turn off the auto-off function, touch the AUTO-OFF button again. The words “AUTO-OFF” will no longer be illuminated in the display. Remember, when AUTO-OFF is on, the unit will automatically turn off the heater when the timer counts down to zero. The user does not have to be present to turn off the heater when it is in this mode.

XI. TEMPERATURE CALIBRATION

The HP30 is designed for accuracy. The temperature calibration designed into the units is made to hold for very long periods of time. When calibrated in the factory, it is expected that the units will meet the most demanding customers' requirements. However, our standards for temperature measurement may not be the same as the users. Therefore, the HS30 has been designed to be calibrated in the field by the user. Follow the instructions below if calibration is wanted or needed.

Note: Since the units come calibrated from the factory, changes to any individual calibration point or points can be made in the field. These points are listed in the procedure below. To recalibrate the entire unit the user may wish to remove all previously entered calibration. To clear the old calibration points , hold the AUTO-OFF key down for three seconds. The unit will beep. Then press and hold either the PLATE or PROBE key for 3 seconds to clear that setting. Repeat the procedure to clear the other if needed. This will erase the calibration data stored in memory.

Follow the instructions below if calibration is wanted or needed.

<u>PROBE CALIBRATION</u>

Probe calibration is performed by using an accessory calibration kit (HS30-700) which precisely simulates fixed temperature points. The kit has two "dummy" probes which, when plugged into the probe jack, represent the value of resistance equal to the temperature shown on the dummy probe. The dummy probes represent temperatures of 25°C and 400°C. These are the probe calibration points. The kit is available from the manufacturer.

To calibrate the probe temperature, follow these steps.

1. Insert the 25°C dummy probe in the probe jack at the rear of the unit. Depress and hold the AUTO-OFF key for three (3) seconds followed by the PROBE key and the probe display will flash. Use the UP or DOWN arrow as needed to set the display to read 25°C. Then press and release the PROBE key again. When the display stops flashing it will read 25°C and the low calibration point has been set.
2. Remove the 25°C dummy probe and replace it with the 400°C dummy probe. Depress and hold the AUTO-OFF key for three (3) seconds followed PROBE KEY and the probe display will flash. Use the UP or DOWN arrow as needed to make the probe display read 400°C. Then press and release the PROBE key again. When the display stops flashing it will read 400°C and the high calibration point has been set.
3. Remove the dummy probe. The probe calibration is now complete.

PLATE CALIBRATION

The plate calibration does not affect the probe calibration and vice versa. If you are using the probe temperature control loop only, it is not necessary to calibrate the plate loop. The plate loop has been factory calibrated and is not likely to change.

To calibrate the plate, one needs an accurate temperature meter, preferably digital, and a surface temperature probe. This is a very difficult measurement to make, and not all surface temperature probes do the job well. If you need help, please call the factory. The Plate and Probe calibration kit #HS30-800 is available from the factory and will contain the temperature meter, probe and dummy probes for the probe calibration.

The calibration procedure requires that calibration starts at the low temperatures and goes high. It is time consuming. Be sure before making an adjustment that the plate temperature has stabilized at the point being calibrated. Follow the procedure below.

1. Start at ambient. Using your temperature meter, read the plate surface temperature in the center of the plate. Push and hold the AUTO-OFF key for three (3) seconds followed by the PLATE key. The plate display will flash. Use the UP or DOWN arrows to make the display agree with the measured plate temperature on your meter. Then press and release the PLATE key again. When the display stops flashing the low calibration is set.
2. Set the plate temperature to 400°C. The plate will start to heat. Give the unit ample time to reach and stabilize at 400°C. When the temperature is stable, measure the plate temperature with your meter. Depress and hold the AUTO-OFF key for three (3) seconds followed by the PLATE key and the plate display will flash. Use the UP or DOWN arrow as needed to make the plate temperature display read what is measured on your meter. Then press and release the PLATE key again. When the display stops flashing the high temperature calibration point is set.

The plate surface calibration is now complete.

XII. POWER INTERRUPTION PROTECTION

The unit is provided with power interruption protection. If AC power is interrupted while running a temperature, the unit will go back to the set temperature when the power returns. To let the user know that a power interruption has occurred, the temperature display will be flashing between TARGET(T) and ACTUAL(A) faster than usual by only dwelling for one second on each. Under normal operating conditions, the display will alternate between (T) and (A) dwelling on (A) for three seconds and (T) for one second. To restore the display to normal, touch the DOWN arrow.

XIII. RS232 INTERFACE

The RS232 is available through the 9-pin D-subminiature connector on the rear of the unit. It operates at 2400 Baud, 8 Data Bits, 1 Stop Bit, No Parity. No handshake hardware or software is necessary. It will work well on a Windows terminal program per the settings listed here. All communications settings and queries are done using ASCII characters with carriage return as the terminating character (CR[1]).

Any serial port on your PC can be used. The pin outs follow standards for RS232C. On the connector, pin 5 is ground, pin 3 is RX and pin 2 is TX. We do not provide software for this application, however, several users have found that using Windows Terminal Program, and Visual Basic works well, but any programming language is OK.

Note: To comply with CE and to avoid possible EMI radiation from the RS232 cable, a shielded cable should be used.

Query	Command
Request current plate temperature	a
Request current probe temperature	b
Request current timer	c
Request ramping value	d
Request plate target temperature	e
Request probe target temperature	f
Request stirrer speed (rpm)	g

For example: To request current plate temperature one needs to send two characters a + CR
and the unit responds xxx + CR where xxx is the present temperature of the heater plate.

Note: If there is no target plate temperature set into the unit it will send “---” + CR.

Set	Command
Plate target temperature	A
Probe target temperature	B
Timer	C
Ramping	D
Stirrer speed(rpm)	E
Stirrer Off	F
Heater Off	G
Toggle Auto ON/OFF	H

For example: To set the following command string do the following:

To set the plate temperature target to 123°C

A 123 + Cr

To set the timer

C 12:5600 + CR

To set the stirrer speed to 500 rpm

E 500 + CR

Note: If the command string is valid, the unit will respond back with the message "Command OK" + CR. If the command is incorrect the message will be "Command Failed" + CR.

XIV. CLEANING AND MAINTENANCE

This unit is subject to splashes and spills during normal use. Be sure to clean all spills quickly. Wipe spills with a soft cloth or paper towel. If a cleaning solution is necessary, use a mild soap or detergent solution and a soft cloth. Be sure not to use solvents. They could damage the paint, labels or display window on the unit.

A mild abrasive can be used to clean the glass heater surface. Do not use abrasives on the cast aluminum top. It will scratch the surface.

<u>CAUTION:</u> DO NOT ATTEMPT TO CLEAN THE HEATER SURFACE WHEN IT IS HOT. BURNS COULD OCCUR.